

Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB137183
Analysis Date: 09/15/2025
BalanceID: WC SC-5
OvenID: EXT OVEN-3

ANALYST: JIGNESH
REVIEWED BY: Iwona
Extraction Date: 09/15/2025
Extraction IN Time: 08:45
Extraction OUT Time: 09:15
Thermometer ID: EXT OVEN#3

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB137183BL	LB137183BL	WATER	1.3	1000	100	3.1526	3.1526	3.02	3.1527	3.1527	0.0001	0.1
2	LB137183BS	LB137183BS	WATER	1.3	1000	100	2.9741	2.9741	3.01	2.9912	2.9912	0.0171	17.1
3	LB137183BSD	LB137183BSD	WATER	1.3	1000	100	3.1998	3.1998	3.03	3.2167	3.2167	0.0169	16.9
4	Q3067-01	GRAB	WATER	1.6	1000	100	3.0737	3.0737	3.04	3.0942	3.0942	0.0205	20.5
5	Q3070-01	DRAIN-WATER-TANK-1	WATER	1.6	1000	100	3.0460	3.0460	3.03	3.0500	3.0500	0.0040	4
6	Q3081-01	DRAIN-WATER-TANK-1	WATER	1.6	1000	100	3.0563	3.0563	3.05	3.0569	3.0569	0.0006	0.6
7	Q3093-01	MH-9-12-25	WATER	1.6	1000	100	3.0107	3.0107	3.04	3.0389	3.0389	0.0282	28.2
8	Q3096-01	DRAIN-WATER-TANK-1	WATER	1.6	1000	100	3.0374	3.0374	3.03	3.0395	3.0395	0.0021	2.1

QC Batch# LB137183

Test: TPH

Analysis Date: 09/15/2025

Chemicals Used:

Chemical Name	Chemical Lot #
HEXANE	W3204
pH Paper 0-14	M6069
Sodium Sulfate	EP2636
1:1 HCL	WP112782
Silica Gel	W3079
Sand	NA

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	5.00 ML	WP112783
LCSWD	5.00 ML	WP112784
MS/MSD	N/A	N/A

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-6

Before Analysis

0.0020 gram Balance: 0.0018 (0.0018-0.0022) In OVEN TEMP1 : 71 °C Dessicator Time In1 : 10:31

1.0000 gram Balance: 1.0004 (0.9950-1.0050) In Time1: 09:40

Bal Check Time: 09:00 Out OVEN TEMP1: 71 °C Dessicator Time Out1: 11:10

Out Time1: 10:30

After Analysis

0.0020 gram Balance: 0.002 (0.0018-0.0022) In OVEN TEMP2 : 70 °C Dessicator Time In2 : 12:32

1.0000 gram Balance: 1.0003 (0.9950-1.0050) In Time2: 12:00

Bal Check Time: 13:15 Out OVEN TEMP2: 70 °C Dessicator Time Out2: 13:10

Out Time2: 12:30

13183

WORKLIST(Hardcopy Internal Chain)

WorkList Name : tph q3096 WorkList ID : 191853 Department : Wet-Chemistry Date : 09-15-2025 07:51:47

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3067-01	GRAB	Water	TPH	Conc H2SO4 to pH < 2	ARAM01	D31	09/10/2025	1664A
Q3070-01	F DRAIN-WATER-TANK-1	Water	TPH	Conc H2SO4 to pH < 2	MAJO01	D21	09/10/2025	1664A
Q3081-01	F DRAIN-WATER-TANK-1	Water	TPH	Conc H2SO4 to pH < 2	MAJO01	D11	09/11/2025	1664A
Q3093-01	F MH-9-12-25	Water	TPH	Conc H2SO4 to pH < 2	EURO03	D31	09/12/2025	1664A
Q3096-01	F DRAIN-WATER-TANK-1	Water	TPH	Conc H2SO4 to pH < 2	MAJO01	D31	09/12/2025	1664A

Date/Time 09-15-25 08:00
Raw Sample Received by: J8 WWC
Raw Sample Relinquished by: J8 WWC

Date/Time 09-15-25 15:00
Raw Sample Received by: J8 WWC
Raw Sample Relinquished by: J8 WWC