

Laboratory Control Sample Summary

Client: Major Products

SDG No.: Q3096

Project: Floor Drain Water 2025

Run No.: LB137180

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137180BS							
BOD5	mg/L	198	181		91	1	84.6-115.4	09/12/2025

Laboratory Control Sample Summary

Client: Major Products

SDG No.: Q3096

Project: Floor Drain Water 2025

Run No.: LB137182

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137182BS							
Oil and Grease	mg/L	20.0	16.8		84	1	78-114	09/15/2025

Laboratory Control Sample Summary

Client: Major Products

SDG No.: Q3096

Project: Floor Drain Water 2025

Run No.: LB137183

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137183BS							
TPH	mg/L	20.0	17.1		86	1	78-114	09/15/2025

Laboratory Control Sample Summary

Client: Major Products

SDG No.: Q3096

Project: Floor Drain Water 2025

Run No.: LB137183

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137183BSD							
TPH	mg/L	20.0	16.9		84	1	78-114	09/15/2025

Laboratory Control Sample Summary

Client: Major Products

SDG No.: Q3096

Project: Floor Drain Water 2025

Run No.: LB137186

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137186BS							
TSS	mg/L	550	531		96	1	90-110	09/15/2025