



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

Matrix Spike/Matrix Spike Duplicate Summary  
SW-846

|          |                                |                    |                 |
|----------|--------------------------------|--------------------|-----------------|
| SDG No.: | Q3097                          | Analytical Method: | SW8270-Modified |
| Client:  | AECOM Technical Services, Inc. | DataFile:          | BN037773.D      |

| Parameter      | Spike       | Sample Result     | Result              | Units | Rec | Rec Qual | RPD | RPD Qual | Low | Limits High | RPD |
|----------------|-------------|-------------------|---------------------|-------|-----|----------|-----|----------|-----|-------------|-----|
| Lab Sample ID: | Q3097-03MSD | Client Sample ID: | RE103D2-20250911MSD |       |     |          |     |          |     |             |     |
| 1,4-Dioxane    | 0.4         | 0.37              | 0.73                | ug/L  | 90  |          | 2   |          | 70  | 130         | 20  |