



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

## Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3097

Analytical Method: 8270-Modified

Client: AECOM Technical Services, Inc.

DataFile: BN037885.D

| Lab Sample ID | Parameter   | Spike | Result | Unit | Rec | RPD | Qual | RPD  | Limits |      |
|---------------|-------------|-------|--------|------|-----|-----|------|------|--------|------|
|               |             |       |        |      |     |     |      | Qual | Low    | High |
| PB169712BS    | 1,4-Dioxane | 0.4   | 0.32   | ug/L | 80  |     |      |      | 70     | 130  |