

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3097

Client ID : SSTDCCC0.4

Date Analyzed: 09/18/2025

Lab File ID: BN037759.D

Time Analyzed: 02:46

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	8012	7.833	21664	10.64	10192	14.48
UPPER LIMIT	16024	8.333	43328	11.137	20384	14.984
LOWER LIMIT	4006	7.333	10832	10.137	5096	13.984
EPA SAMPLE NO.						
01 RE125D2-20250911	6450	7.83	17916	10.64	8773	14.48
02 RE103D2-20250911	8062	7.83	22030	10.64	10128	14.48
03 RE103D2-20250911MS	7218	7.83	19437	10.64	8792	14.48
04 RE103D2-20250911MSD	7159	7.83	19099	10.64	8494	14.48
05 RE108D2-20250911	7633	7.83	20699	10.64	9180	14.48

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.