

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3097

Client ID : SSTDCCC0.4

Date Analyzed: 09/20/2025

Lab File ID: BN037832.D

Time Analyzed: 02:35

Instrument ID: BNA\_N

GC Column: ZB-GR ID: 0.25 (mm)

|                   | IS1 (DCB)<br>AREA # | RT #  | IS2 (NPT)<br>AREA # | RT #   | IS3 (ANT)<br>AREA # | RT #   |
|-------------------|---------------------|-------|---------------------|--------|---------------------|--------|
| 12 HOUR STD       | 6232                | 7.825 | 16500               | 10.64  | 7183                | 14.48  |
| UPPER LIMIT       | 12464               | 8.325 | 33000               | 11.137 | 14366               | 14.984 |
| LOWER LIMIT       | 3116                | 7.325 | 8250                | 10.137 | 3591.5              | 13.984 |
| EPA SAMPLE NO.    |                     |       |                     |        |                     |        |
| 01 DUP06-20250912 | 5678                | 7.83  | 14769               | 10.63  | 5979                | 14.48  |

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.

\* Values outside of QC limits.