

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3097

Client ID : SSTDCCC0.4

Date Analyzed: 09/22/2025

Lab File ID: BN037849.D

Time Analyzed: 10:19

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	6007	7.833	15935	10.64	7024	14.48
UPPER LIMIT	12014	8.333	31870	11.137	14048	14.984
LOWER LIMIT	3003.5	7.333	7967.5	10.137	3512	13.984
EPA SAMPLE NO.						
01 DUP06-20250912DL	5114	7.83	13234	10.63	5436	14.47
02 RE134D2-20250912DL	5950	7.83	15397	10.63	6317	14.48
03 RE134D3-20250912DL	6180	7.83	16337	10.63	6729	14.48

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.