

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3097

Client ID : SSTDCCC0.4

Date Analyzed: 09/25/2025

Lab File ID: BN037883.D

Time Analyzed: 11:31

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	5686	7.818	16943	10.63	8938	14.47
UPPER LIMIT	11372	8.318	33886	11.126	17876	14.973
LOWER LIMIT	2843	7.318	8471.5	10.126	4469	13.973
EPA SAMPLE NO.						
01 PB169712BS	7082	7.82	19552	10.63	9660	14.47
02 PB169712BSD	6836	7.83	18512	10.63	8848	14.47

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.