

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BN037758.D
Instrument ID: BNA_N

Contract: AECO15
SDG NO.: Q3097
DFTPP Injection Date: 09/18/2025
DFTPP Injection Time: 02:06

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.4 (1.1) 1
69	Mass 69 relative abundance	100
70	Less than 2.0% of mass 69	0.2 (0.5) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.9
365	Greater than 1% of mass 198	3.3
441	Present, but less than mass 443	86.6
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	16.5 (19.9) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN037759.D	09/18/2025	02:46
RE108D2-20250911	Q3097-07	BN037769.D	09/18/2025	08:47
RE125D2-20250911	Q3097-08	BN037770.D	09/18/2025	09:23
RE103D2-20250911	Q3097-01	BN037771.D	09/18/2025	10:00
RE103D2-20250911MS	Q3097-02MS	BN037772.D	09/18/2025	11:48
RE103D2-20250911MSD	Q3097-03MSD	BN037773.D	09/18/2025	12:25
SSTDCCC0.4EC	SSTDCCC0.4	BN037774.D	09/18/2025	13:16