

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>AECO15</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3097</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>09/22/2025 19:40</u>
Lab File ID:	<u>VX047717.D</u>	Init. Calib. Date(s):	<u>09/16/2025 09/16/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:23 12:01</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.518	0.586		13.13	50
Chloromethane	0.680	0.730	0.1	7.35	50
Vinyl Chloride	0.666	0.709		6.46	50
Bromomethane	0.422	0.480		13.74	50
Chloroethane	0.445	0.467		4.94	50
Trichlorofluoromethane	0.989	0.995		0.61	50
1,1,2-Trichlorotrifluoroethane	0.578	0.585		1.21	50
1,1-Dichloroethene	0.594	0.603		1.51	50
Acetone	0.346	0.287		-17.05	50
Carbon Disulfide	1.626	1.720		5.78	50
Methyl tert-butyl Ether	2.169	2.272		4.75	50
Methyl Acetate	0.770	0.852		10.65	50
Methylene Chloride	0.732	0.748		2.19	50
trans-1,2-Dichloroethene	0.640	0.668		4.38	50
1,1-Dichloroethane	1.263	1.307	0.1	3.48	50
Cyclohexane	1.052	1.042		-0.95	50
2-Butanone	0.432	0.447		3.47	50
Carbon Tetrachloride	0.532	0.503		-5.45	50
cis-1,2-Dichloroethene	0.783	0.813		3.83	50
Bromochloromethane	0.551	0.567		2.9	50
Chloroform	1.276	1.327		4	50
1,1,1-Trichloroethane	1.082	1.106		2.22	50
Methylcyclohexane	0.556	0.540		-2.88	50
Benzene	1.488	1.513		1.68	50
1,2-Dichloroethane	0.566	0.561		-0.88	50
Trichloroethene	0.360	0.374		3.89	50
1,2-Dichloropropane	0.381	0.393		3.15	50
Bromodichloromethane	0.572	0.580		1.4	50
4-Methyl-2-Pentanone	0.477	0.520		9.02	50
Toluene	0.917	0.927		1.09	50
t-1,3-Dichloropropene	0.584	0.596		2.06	50
cis-1,3-Dichloropropene	0.624	0.637		2.08	50
1,1,2-Trichloroethane	0.354	0.369		4.24	50
2-Hexanone	0.343	0.363		5.83	50
Dibromochloromethane	0.407	0.428		5.16	50
1,2-Dibromoethane	0.360	0.387		7.5	50
Tetrachloroethene	0.326	0.319		-2.15	50
Chlorobenzene	1.136	1.142	0.3	0.53	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>AECO15</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3097</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>09/22/2025 19:40</u>
Lab File ID:	<u>VX047717.D</u>	Init. Calib. Date(s):	<u>09/16/2025 09/16/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:23 12:01</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.960	1.952		-0.41	50
m/p-Xylenes	0.729	0.736		0.96	50
o-Xylene	0.706	0.717		1.56	50
Styrene	1.236	1.262		2.1	50
Bromoform	0.293	0.301	0.1	2.73	50
Isopropylbenzene	3.801	3.740		-1.61	50
1,1,2,2-Tetrachloroethane	1.150	1.212	0.3	5.39	50
1,3-Dichlorobenzene	1.739	1.722		-0.98	50
1,4-Dichlorobenzene	1.785	1.742		-2.41	50
1,2-Dichlorobenzene	1.657	1.679		1.33	50
1,2-Dibromo-3-Chloropropane	0.233	0.247		6.01	50
1,2,4-Trichlorobenzene	1.077	1.057		-1.86	50
1,2,3-Trichlorobenzene	1.002	1.015		1.3	50
1,2-Dichloroethane-d4	0.796	0.871		9.42	50
Dibromofluoromethane	0.335	0.370		10.45	50
Toluene-d8	1.160	1.276		10	50
4-Bromofluorobenzene	0.440	0.489		11.14	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.