

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>AECO15</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3097</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>09/23/2025</u> <u>09:02</u>
Lab File ID:	<u>VX047719.D</u>	Init. Calib. Date(s):	<u>09/16/2025</u> <u>09/16/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:23</u> <u>12:01</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.518	0.656		26.64	20
Chloromethane	0.680	0.808	0.1	18.82	20
Vinyl Chloride	0.666	0.805		20.87	20
Bromomethane	0.422	0.530		25.59	20
Chloroethane	0.445	0.514		15.51	20
Trichlorofluoromethane	0.989	1.104		11.63	20
1,1,2-Trichlorotrifluoroethane	0.578	0.650		12.46	20
1,1-Dichloroethene	0.594	0.677		13.97	20
Acetone	0.346	0.388		12.14	20
Carbon Disulfide	1.626	1.899		16.79	20
Methyl tert-butyl Ether	2.169	2.452		13.05	20
Methyl Acetate	0.770	0.873		13.38	20
Methylene Chloride	0.732	0.832		13.66	20
trans-1,2-Dichloroethene	0.640	0.741		15.78	20
1,1-Dichloroethane	1.263	1.428	0.1	13.06	20
Cyclohexane	1.052	1.142		8.56	20
2-Butanone	0.432	0.477		10.42	20
Carbon Tetrachloride	0.532	0.561		5.45	20
cis-1,2-Dichloroethene	0.783	0.903		15.33	20
Bromochloromethane	0.551	0.612		11.07	20
Chloroform	1.276	1.444		13.17	20
1,1,1-Trichloroethane	1.082	1.211		11.92	20
Methylcyclohexane	0.556	0.599		7.73	20
Benzene	1.488	1.647		10.69	20
1,2-Dichloroethane	0.566	0.606		7.07	20
Trichloroethene	0.360	0.403		11.94	20
1,2-Dichloropropane	0.381	0.426		11.81	20
Bromodichloromethane	0.572	0.634		10.84	20
4-Methyl-2-Pentanone	0.477	0.525		10.06	20
Toluene	0.917	1.016		10.8	20
t-1,3-Dichloropropene	0.584	0.658		12.67	20
cis-1,3-Dichloropropene	0.624	0.704		12.82	20
1,1,2-Trichloroethane	0.354	0.392		10.73	20
2-Hexanone	0.343	0.374		9.04	20
Dibromochloromethane	0.407	0.464		14.01	20
1,2-Dibromoethane	0.360	0.410		13.89	20
Tetrachloroethene	0.326	0.353		8.28	20
Chlorobenzene	1.136	1.246	0.3	9.68	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>AECO15</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3097</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>09/23/2025 09:02</u>
Lab File ID:	<u>VX047719.D</u>	Init. Calib. Date(s):	<u>09/16/2025 09/16/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:23 12:01</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.960	2.150		9.69	20
m/p-Xylenes	0.729	0.807		10.7	20
o-Xylene	0.706	0.790		11.9	20
Styrene	1.236	1.382		11.81	20
Bromoform	0.293	0.332	0.1	13.31	20
Isopropylbenzene	3.801	4.072		7.13	20
1,1,2,2-Tetrachloroethane	1.150	1.253	0.3	8.96	20
1,3-Dichlorobenzene	1.739	1.887		8.51	20
1,4-Dichlorobenzene	1.785	1.893		6.05	20
1,2-Dichlorobenzene	1.657	1.825		10.14	20
1,2-Dibromo-3-Chloropropane	0.233	0.241		3.43	20
1,2,4-Trichlorobenzene	1.077	1.150		6.78	20
1,2,3-Trichlorobenzene	1.002	1.103		10.08	20
1,2-Dichloroethane-d4	0.796	0.816		2.51	20
Dibromofluoromethane	0.335	0.352		5.07	20
Toluene-d8	1.160	1.200		3.45	20
4-Bromofluorobenzene	0.440	0.459		4.32	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.