

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>AECO15</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3097</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>09/23/2025 19:00</u>
Lab File ID:	<u>VX047744.D</u>	Init. Calib. Date(s):	<u>09/16/2025 09/16/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:23 12:01</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.518	0.632		22.01	50
Chloromethane	0.680	0.784	0.1	15.29	50
Vinyl Chloride	0.666	0.776		16.52	50
Bromomethane	0.422	0.496		17.54	50
Chloroethane	0.445	0.505		13.48	50
Trichlorofluoromethane	0.989	1.097		10.92	50
1,1,2-Trichlorotrifluoroethane	0.578	0.639		10.55	50
1,1-Dichloroethene	0.594	0.657		10.61	50
Acetone	0.346	0.295		-14.74	50
Carbon Disulfide	1.626	1.828		12.42	50
Methyl tert-butyl Ether	2.169	2.338		7.79	50
Methyl Acetate	0.770	0.876		13.77	50
Methylene Chloride	0.732	0.797		8.88	50
trans-1,2-Dichloroethene	0.640	0.715		11.72	50
1,1-Dichloroethane	1.263	1.376	0.1	8.95	50
Cyclohexane	1.052	1.100		4.56	50
2-Butanone	0.432	0.442		2.32	50
Carbon Tetrachloride	0.532	0.522		-1.88	50
cis-1,2-Dichloroethene	0.783	0.862		10.09	50
Bromochloromethane	0.551	0.579		5.08	50
Chloroform	1.276	1.391		9.01	50
1,1,1-Trichloroethane	1.082	1.150		6.28	50
Methylcyclohexane	0.556	0.569		2.34	50
Benzene	1.488	1.586		6.59	50
1,2-Dichloroethane	0.566	0.582		2.83	50
Trichloroethene	0.360	0.386		7.22	50
1,2-Dichloropropane	0.381	0.410		7.61	50
Bromodichloromethane	0.572	0.607		6.12	50
4-Methyl-2-Pentanone	0.477	0.521		9.22	50
Toluene	0.917	0.974		6.22	50
t-1,3-Dichloropropene	0.584	0.590		1.03	50
cis-1,3-Dichloropropene	0.624	0.642		2.88	50
1,1,2-Trichloroethane	0.354	0.387		9.32	50
2-Hexanone	0.343	0.358		4.37	50
Dibromochloromethane	0.407	0.441		8.35	50
1,2-Dibromoethane	0.360	0.396		10	50
Tetrachloroethene	0.326	0.337		3.37	50
Chlorobenzene	1.136	1.185	0.3	4.31	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>AECO15</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3097</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>09/23/2025 19:00</u>
Lab File ID:	<u>VX047744.D</u>	Init. Calib. Date(s):	<u>09/16/2025 09/16/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:23 12:01</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.960	2.020		3.06	50
m/p-Xylenes	0.729	0.755		3.57	50
o-Xylene	0.706	0.741		4.96	50
Styrene	1.236	1.305		5.58	50
Bromoform	0.293	0.305	0.1	4.1	50
Isopropylbenzene	3.801	3.780		-0.55	50
1,1,2,2-Tetrachloroethane	1.150	1.194	0.3	3.83	50
1,3-Dichlorobenzene	1.739	1.714		-1.44	50
1,4-Dichlorobenzene	1.785	1.729		-3.14	50
1,2-Dichlorobenzene	1.657	1.655		-0.12	50
1,2-Dibromo-3-Chloropropane	0.233	0.228		-2.15	50
1,2,4-Trichlorobenzene	1.077	1.029		-4.46	50
1,2,3-Trichlorobenzene	1.002	0.989		-1.3	50
1,2-Dichloroethane-d4	0.796	0.781		-1.88	50
Dibromofluoromethane	0.335	0.336		0.3	50
Toluene-d8	1.160	1.139		-1.81	50
4-Bromofluorobenzene	0.440	0.434		-1.36	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.