

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: AECO15
 Lab Code: ACE SDG NO.: Q3097
 Lab File ID: VX047660.D Date Analyzed: 09/19/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:22
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	129778	12.00€				
UPPER LIMIT	259556	12.50€				
LOWER LIMIT	64889	11.50€				
EPA SAMPLE NO.						
RE125D1-20250911	146768	12.01				
RE125D3-20250911	143358	12.01				
RE103D3-20250911	134673	12.01				
RE108D2-20250911	137704	12.01				
RE125D2-20250911	144012	12.01				
RE105D1-20250911	137612	12.01				
DUP04-20250911	140501	12.01				
DUP05-20250911	146680	12.01				
RE105D2-20250911	134519	12.01				
VX0919WBL01	175228	12.01				
VX0919WBS01	146041	12.01				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.