

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>AECO15</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3097</u>
Lab File ID:	<u>VX047689.D</u>	Date Analyzed:	<u>09/22/2025</u>
Instrument ID:	<u>MSVOA_X</u>	Time Analyzed:	<u>09:38</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)
		Heated Purge:	(Y/N) <u>N</u>

	IS4 AREA #	RT #				
12 HOUR STD	135594	12.00€				
UPPER LIMIT	271188	12.50€				
LOWER LIMIT	67797	11.50€				
EPA SAMPLE NO.						
RE103D2-20250911	156017	12.01				
RE103D2-20250911MS	116537	12.01				
RE103D2-20250911MSD	125295	12.01				
RE108D2-20250911DL	168884	12.01				
RE125D2-20250911DL	163574	12.01				
RE108D1-20250911	152363	12.01				
RE120D3-20250911	149631	12.01				
RE105D2-20250911DL	159199	12.01				
EB01-20250912	150335	12.01				
VX0922WBL01	164144	12.01				
VX0922WBS01	141526	12.00				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.