

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: AECO15
 Lab Code: ACE SDG NO.: Q3097
 Lab File ID: VX047719.D Date Analyzed: 09/23/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:02
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	115061	12.00€				
UPPER LIMIT	230122	12.50€				
LOWER LIMIT	57530.5	11.50€				
EPA SAMPLE NO.						
RE103D2-20250911DL	132316	12.01				
RE117D2-20250911	143533	12.01				
TB03-20250911	146254	12.01				
TB04-20250911	135735	12.01				
TB05-20250912	151284	12.01				
RE120D1-20250912	141902	12.01				
DUP06-20250912	141279	12.01				
RE120D2-20250912	131853	12.01				
RE134D1-20250912	139922	12.01				
RE134D4-20250912	144658	12.01				
RE134D2-20250912	155005	12.01				
VX0923WBL01	132951	12.01				
VX0923WBS02	141661	12.01				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.