

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: AECO15  
 Lab Code: ACE SDG NO.: Q3097  
 Lab File ID: VX047746.D Date Analyzed: 09/24/2025  
 Instrument ID: MSVOA\_X Time Analyzed: 02:08  
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

|                  | IS4<br>AREA # | RT #   |  |  |  |  |
|------------------|---------------|--------|--|--|--|--|
| 12 HOUR STD      | 121948        | 12.00€ |  |  |  |  |
| UPPER LIMIT      | 243896        | 12.50€ |  |  |  |  |
| LOWER LIMIT      | 60974         | 11.50€ |  |  |  |  |
| EPA SAMPLE NO.   |               |        |  |  |  |  |
| RE134D3-20250912 | 134755        | 12.01  |  |  |  |  |
| VX0923WBL02      | 171478        | 12.01  |  |  |  |  |
| VX0923WBS03      | 123765        | 12.01  |  |  |  |  |

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.