

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: AECO15
 Lab Code: ACE SDG NO.: Q3097
 Lab File ID: VX047769.D Date Analyzed: 09/24/2025
 Instrument ID: MSVOA_X Time Analyzed: 11:46
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	114093	12.00€				
UPPER LIMIT	228186	12.50€				
LOWER LIMIT	57046.5	11.50€				
EPA SAMPLE NO.						
RE120D1-20250912DL	140354	12.01				
DUP06-20250912DL	149875	12.01				
RE120D2-20250912DL	133342	12.01				
RE134D4-20250912DL	148581	12.01				
RE134D3-20250912DL	143934	12.01				
VX0924WBL01	123109	12.01				
VX0924WBS01	129236	12.01				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.