

Laboratory Control Sample Summary

Client: Ardmore Chemical

SDG No.: Q3098

Project: PVSC Monthly 2025

Run No.: LB137180

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137180BS							
BOD5	mg/L	198	181		91	1	84.6-115.4	09/12/2025

Laboratory Control Sample Summary

Client: Ardmore Chemical

SDG No.: Q3098

Project: PVSC Monthly 2025

Run No.: LB137186

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137186BS							
TSS	mg/L	550	531		96	1	90-110	09/15/2025

Laboratory Control Sample Summary

Client: Ardmore Chemical

SDG No.: Q3098

Project: PVSC Monthly 2025

Run No.: LB137201

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB169643BS							
Cyanide	mg/L	0.1	0.096		96	1	85-115	09/16/2025