

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3098

Client ID : SSTDCCC040

Date Analyzed: 09/18/2025

Lab File ID: BP025785.D

Time Analyzed: 17:24

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	161596	7.755	642758	10.52	424090	14.38
UPPER LIMIT	323192	8.255	1285520	11.019	848180	14.878
LOWER LIMIT	80798	7.255	321379	10.019	212045	13.878
EPA SAMPLE NO.						
01 PB169718BL	153939	7.75	582617	10.52	356809	14.38
02 PB169718BS	175687	7.75	683598	10.52	444972	14.38

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.