

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3098

Client ID : SSTDCCC040

Date Analyzed: 09/19/2025

Lab File ID: BP025792.D

Time Analyzed: 10:12

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	182285	7.749	709982	10.51	456027	14.37
UPPER LIMIT	364570	8.249	1419960	11.013	912054	14.866
LOWER LIMIT	91142.5	7.249	354991	10.013	228014	13.866
EPA SAMPLE NO.						
01 EFF-WW	178989	7.75	525343	10.52	428315	14.37
02 EFF-WWMS	174041	7.75	537202	10.52	439134	14.38
03 EFF-WWMSD	184711	7.75	565127	10.52	444962	14.39

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.