

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BP025791.D
Instrument ID: BNA_P

Contract: ARDM01
SDG NO.: Q3098
DFTPP Injection Date: 09/19/2025
DFTPP Injection Time: 09:31

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.6 (1.6) 1
69	Mass 69 relative abundance	37.8
70	Less than 2.0% of mass 69	0.2 (0.5) 1
197	Less than 2.0% of mass 198	0.4
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.9
365	Greater than 1% of mass 198	3.4
441	Present, but less than mass 443	10.3
442	Greater than 50% of mass 198	66.8
443	15.0 - 24.0% of mass 442	13 (19.5) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BP025792.D	09/19/2025	10:12
EFF-WW	Q3098-04	BP025801.D	09/19/2025	16:28
EFF-WWMS	Q3098-05MS	BP025802.D	09/19/2025	17:09
EFF-WWMSD	Q3098-06MSD	BP025803.D	09/19/2025	17:51