

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance

Contract: ARDM01

Lab Code: ACE

SDG No.: Q3098

Instrument ID: MSVOA_N

Calibration Date(s): 09/04/2025 09/04/2025

Heated Purge: (Y/N) N

Calibration Time(s): 16:23 17:46

GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:		RRF005 = VN087725.D		RRF020 = VN087726.D		RRF050 = VN087727.D		RRF100 = VN087728.D		RRF150 = VN087729.D		RRF =			
COMPOUND		RRF005	RRF020	RRF050	RRF100	RRF150	RRF								% RSD
Chloromethane		1.939	1.792	2.026	2.037	1.992		1.957							5.1
Vinyl Chloride		2.110	2.001	2.388	2.315	2.249		2.213							7.1
Bromomethane		1.293	1.108	1.325	1.377	1.387		1.298							8.7
Chloroethane		1.599	1.358	1.577	1.521	1.506		1.512							6.2
Trichlorofluoromethane		3.241	2.904	3.445	3.355	3.271		3.243							6.3
1,1-Dichloroethene		1.836	1.575	1.915	1.883	1.846		1.811							7.5
Acrolein		0.523	0.431	0.496	0.518	0.518		0.497							7.7
Acrylonitrile		1.167	1.051	1.253	1.274	1.263		1.202							7.9
Methylene Chloride		2.174	1.759	2.150	2.090	2.084		2.052							8.2
trans-1,2-Dichloroethene		1.709	1.652	1.947	1.965	1.962		1.847							8.3
1,1-Dichloroethane		3.750	3.277	3.899	3.878	3.823		3.725							6.9
Carbon Tetrachloride		0.526	0.466	0.530	0.533	0.522		0.515							5.5
Chloroform		3.785	3.484	4.143	4.049	4.035		3.899							6.9
1,1,1-Trichloroethane		0.599	0.582	0.635	0.647	0.628		0.618							4.4
Benzene		1.514	1.442	1.630	1.657	1.636		1.576							5.9
1,2-Dichloroethane		0.595	0.554	0.609	0.618	0.600		0.595							4.1
Trichloroethene		0.352	0.323	0.364	0.366	0.357		0.352							5
1,2-Dichloropropane		0.394	0.374	0.407	0.428	0.415		0.404							5
Bromodichloromethane		0.612	0.570	0.616	0.647	0.624		0.614							4.5
Toluene		1.705	1.673	1.868	1.935	1.893		1.815							6.5
t-1,3-Dichloropropene		0.594	0.556	0.653	0.696	0.689		0.638							9.5
cis-1,3-Dichloropropene		0.607	0.566	0.678	0.712	0.702		0.653							9.7
1,1,2-Trichloroethane		0.388	0.354	0.399	0.401	0.398		0.388							5.1
2-Chloroethyl vinyl ether		0.310	0.319	0.339	0.377	0.384		0.346							9.7
Dibromochloromethane		0.420	0.369	0.454	0.468	0.464		0.435							9.5
Tetrachloroethene		0.312	0.279	0.297	0.306	0.300		0.299							4.2
Chlorobenzene		1.063	1.006	1.127	1.191	1.182		1.114							7.1
Ethyl Benzene		1.736	1.744	2.042	2.145	2.122		1.958							10.3
m/p-Xylenes		0.629	0.642	0.776	0.800	0.785		0.727							11.5
o-Xylene		0.603	0.639	0.731	0.774	0.776		0.705							11.2

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: <u>Alliance</u>	Contract: <u>ARDM01</u>
Lab Code: <u>ACE</u>	SDG No.: <u>Q3098</u>
Instrument ID: <u>MSVOA_N</u>	Calibration Date(s): <u>09/04/2025</u> <u>09/04/2025</u>
Heated Purge: (Y/N) <u>N</u>	Calibration Time(s): <u>16:23</u> <u>17:46</u>
GC Column: <u>RXI-624</u> ID: <u>0.25</u> (mm)	

LAB FILE ID:		RRF005 = VN087725.D		RRF020 = VN087726.D		RRF050 = VN087727.D	
		RRF100 = VN087728.D		RRF150 = VN087729.D		RRF =	
COMPOUND	RRF005	RRF020	RRF050	RRF100	RRF150	RRF	% RSD
Bromoform	0.265	0.242	0.289	0.312	0.313	0.284	10.8
1,1,2,2-Tetrachloroethane	0.616	0.605	0.666	0.670	0.655	0.642	4.6
1,3-Dichlorobenzene	0.762	0.789	0.888	0.900	0.874	0.843	7.5
1,4-Dichlorobenzene	0.778	0.776	0.895	0.908	0.867	0.845	7.5
1,2-Dichlorobenzene	0.718	0.733	0.839	0.855	0.828	0.795	8.1
1,2-Dichloroethane-d4	2.618	2.509	2.604	2.499	2.583	2.563	2.2
Toluene-d8	1.403	1.411	1.392	1.402	1.369	1.395	1.2
4-Bromofluorobenzene	0.454	0.489	0.475	0.501	0.492	0.482	3.8

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.