

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ATCE02

Lab Code: ACE

SDG NO.: Q3099

Continuing Calib Date: 09/16/2025

Initial Calibration Date(s): 09/15/2025 09/15/2025

Continuing Calib Time: 23:32

Initial Calibration Time(s): 09:34 19:21

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1016-1       | (1) | 4.74       | 4.74      | 4.64      | 4.84 | 0.00       |
| Aroclor-1016-2       | (2) | 4.76       | 4.76      | 4.66      | 4.86 | 0.00       |
| Aroclor-1016-3       | (3) | 4.82       | 4.82      | 4.72      | 4.92 | 0.00       |
| Aroclor-1016-4       | (4) | 4.94       | 4.94      | 4.84      | 5.04 | 0.00       |
| Aroclor-1016-5       | (5) | 5.19       | 5.19      | 5.09      | 5.29 | 0.00       |
| Aroclor-1260-1       | (1) | 6.23       | 6.23      | 6.13      | 6.33 | 0.00       |
| Aroclor-1260-2       | (2) | 6.42       | 6.42      | 6.32      | 6.52 | 0.00       |
| Aroclor-1260-3       | (3) | 6.78       | 6.78      | 6.68      | 6.88 | 0.00       |
| Aroclor-1260-4       | (4) | 7.04       | 7.04      | 6.94      | 7.14 | 0.00       |
| Aroclor-1260-5       | (5) | 7.29       | 7.29      | 7.19      | 7.39 | 0.01       |
| Tetrachloro-m-xylene |     | 3.66       | 3.66      | 3.56      | 3.76 | 0.00       |
| Decachlorobiphenyl   |     | 8.67       | 8.67      | 8.57      | 8.77 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ATCE02

Lab Code: ACE

SDG NO.: Q3099

Continuing Calib Date: 09/16/2025

Initial Calibration Date(s): 09/15/2025

09/15/2025

Continuing Calib Time: 23:32

Initial Calibration Time(s): 09:34

19:21

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1016-1       | (1) | 4.72       | 4.72      | 4.62      | 4.82 | 0.00       |
| Aroclor-1016-2       | (2) | 4.74       | 4.74      | 4.64      | 4.84 | 0.00       |
| Aroclor-1016-3       | (3) | 4.92       | 4.92      | 4.82      | 5.02 | 0.01       |
| Aroclor-1016-4       | (4) | 4.96       | 4.96      | 4.86      | 5.06 | 0.00       |
| Aroclor-1016-5       | (5) | 5.17       | 5.17      | 5.07      | 5.27 | 0.00       |
| Aroclor-1260-1       | (1) | 6.19       | 6.20      | 6.10      | 6.30 | 0.01       |
| Aroclor-1260-2       | (2) | 6.38       | 6.38      | 6.28      | 6.48 | 0.00       |
| Aroclor-1260-3       | (3) | 6.53       | 6.54      | 6.44      | 6.64 | 0.01       |
| Aroclor-1260-4       | (4) | 7.00       | 7.00      | 6.90      | 7.10 | 0.00       |
| Aroclor-1260-5       | (5) | 7.25       | 7.25      | 7.15      | 7.35 | 0.01       |
| Tetrachloro-m-xylene |     | 3.65       | 3.65      | 3.55      | 3.75 | 0.00       |
| Decachlorobiphenyl   |     | 8.61       | 8.61      | 8.51      | 8.71 | 0.00       |