

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q3099
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/15/2025 09/15/2025

Client Sample No.: CCAL02 **Date Analyzed:** 09/16/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113770.D **Time Analyzed:** 18:06

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.740	4.643	4.843	485.110	500.000	-3.0
Aroclor-1016-2	4.759	4.661	4.861	498.650	500.000	-0.3
Aroclor-1016-3	4.816	4.718	4.918	477.770	500.000	-4.4
Aroclor-1016-4	4.935	4.837	5.037	476.930	500.000	-4.6
Aroclor-1016-5	5.192	5.093	5.293	533.220	500.000	6.6
Aroclor-1260-1	6.226	6.129	6.329	519.710	500.000	3.9
Aroclor-1260-2	6.417	6.319	6.519	454.460	500.000	-9.1
Aroclor-1260-3	6.782	6.683	6.883	437.080	500.000	-12.6
Aroclor-1260-4	7.040	6.941	7.141	601.610	500.000	20.3
Aroclor-1260-5	7.284	7.185	7.385	578.330	500.000	15.7
Decachlorobiphenyl	8.669	8.570	8.770	49.960	50.000	-0.1
Tetrachloro-m-xylene	3.658	3.560	3.760	51.610	50.000	3.2

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q3099
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/15/2025 09/15/2025

Client Sample No.: CCAL02 **Date Analyzed:** 09/16/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113770.D **Time Analyzed:** 18:06

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.722	4.623	4.823	474.420	500.000	-5.1
Aroclor-1016-2	4.740	4.641	4.841	476.810	500.000	-4.6
Aroclor-1016-3	4.914	4.816	5.016	483.100	500.000	-3.4
Aroclor-1016-4	4.957	4.858	5.058	480.380	500.000	-3.9
Aroclor-1016-5	5.169	5.070	5.270	494.190	500.000	-1.2
Aroclor-1260-1	6.194	6.096	6.296	452.150	500.000	-9.6
Aroclor-1260-2	6.383	6.284	6.484	446.070	500.000	-10.8
Aroclor-1260-3	6.533	6.435	6.635	423.140	500.000	-15.4
Aroclor-1260-4	7.002	6.903	7.103	435.240	500.000	-13.0
Aroclor-1260-5	7.245	7.146	7.346	439.940	500.000	-12.0
Decachlorobiphenyl	8.614	8.514	8.714	49.240	50.000	-1.5
Tetrachloro-m-xylene	3.651	3.551	3.751	49.590	50.000	-0.8