

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** ATCE02  
**Lab Code:** ACE **SDG NO.:** Q3100  
**GC Column:** ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/15/2025 09/15/2025

**Client Sample No.:** CCAL02 **Date Analyzed:** 09/16/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PO113770.D **Time Analyzed:** 18:06

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D    |
|----------------------|-------|-----------|-------|--------------------|-------------------|-------|
|                      |       | FROM      | TO    |                    |                   |       |
| Aroclor-1016-1       | 4.740 | 4.643     | 4.843 | 485.110            | 500.000           | -3.0  |
| Aroclor-1016-2       | 4.759 | 4.661     | 4.861 | 498.650            | 500.000           | -0.3  |
| Aroclor-1016-3       | 4.816 | 4.718     | 4.918 | 477.770            | 500.000           | -4.4  |
| Aroclor-1016-4       | 4.935 | 4.837     | 5.037 | 476.930            | 500.000           | -4.6  |
| Aroclor-1016-5       | 5.192 | 5.093     | 5.293 | 533.220            | 500.000           | 6.6   |
| Aroclor-1260-1       | 6.226 | 6.129     | 6.329 | 519.710            | 500.000           | 3.9   |
| Aroclor-1260-2       | 6.417 | 6.319     | 6.519 | 454.460            | 500.000           | -9.1  |
| Aroclor-1260-3       | 6.782 | 6.683     | 6.883 | 437.080            | 500.000           | -12.6 |
| Aroclor-1260-4       | 7.040 | 6.941     | 7.141 | 601.610            | 500.000           | 20.3  |
| Aroclor-1260-5       | 7.284 | 7.185     | 7.385 | 578.330            | 500.000           | 15.7  |
| Decachlorobiphenyl   | 8.669 | 8.570     | 8.770 | 49.960             | 50.000            | -0.1  |
| Tetrachloro-m-xylene | 3.658 | 3.560     | 3.760 | 51.610             | 50.000            | 3.2   |

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** ATCE02  
**Lab Code:** ACE **SDG NO.:** Q3100  
**GC Column:** ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/15/2025 09/15/2025

**Client Sample No.:** CCAL02 **Date Analyzed:** 09/16/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PO113770.D **Time Analyzed:** 18:06

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D    |
|----------------------|-------|-----------|-------|--------------------|-------------------|-------|
|                      |       | FROM      | TO    |                    |                   |       |
| Aroclor-1016-1       | 4.722 | 4.623     | 4.823 | 474.420            | 500.000           | -5.1  |
| Aroclor-1016-2       | 4.740 | 4.641     | 4.841 | 476.810            | 500.000           | -4.6  |
| Aroclor-1016-3       | 4.914 | 4.816     | 5.016 | 483.100            | 500.000           | -3.4  |
| Aroclor-1016-4       | 4.957 | 4.858     | 5.058 | 480.380            | 500.000           | -3.9  |
| Aroclor-1016-5       | 5.169 | 5.070     | 5.270 | 494.190            | 500.000           | -1.2  |
| Aroclor-1260-1       | 6.194 | 6.096     | 6.296 | 452.150            | 500.000           | -9.6  |
| Aroclor-1260-2       | 6.383 | 6.284     | 6.484 | 446.070            | 500.000           | -10.8 |
| Aroclor-1260-3       | 6.533 | 6.435     | 6.635 | 423.140            | 500.000           | -15.4 |
| Aroclor-1260-4       | 7.002 | 6.903     | 7.103 | 435.240            | 500.000           | -13.0 |
| Aroclor-1260-5       | 7.245 | 7.146     | 7.346 | 439.940            | 500.000           | -12.0 |
| Decachlorobiphenyl   | 8.614 | 8.514     | 8.714 | 49.240             | 50.000            | -1.5  |
| Tetrachloro-m-xylene | 3.651 | 3.551     | 3.751 | 49.590             | 50.000            | -0.8  |