

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q3100
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/15/2025 09/15/2025

Client Sample No.: CCAL03 **Date Analyzed:** 09/16/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113784.D **Time Analyzed:** 23:32

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.742	4.643	4.843	495.790	500.000	-0.8
Aroclor-1016-2	4.760	4.661	4.861	507.500	500.000	1.5
Aroclor-1016-3	4.818	4.718	4.918	493.690	500.000	-1.3
Aroclor-1016-4	4.937	4.837	5.037	490.220	500.000	-2.0
Aroclor-1016-5	5.193	5.093	5.293	577.640	500.000	15.5
Aroclor-1260-1	6.229	6.129	6.329	524.100	500.000	4.8
Aroclor-1260-2	6.418	6.319	6.519	508.270	500.000	1.7
Aroclor-1260-3	6.783	6.683	6.883	516.710	500.000	3.3
Aroclor-1260-4	7.041	6.941	7.141	587.260	500.000	17.5
Aroclor-1260-5	7.285	7.185	7.385	595.630	500.000	19.1
Decachlorobiphenyl	8.669	8.570	8.770	54.730	50.000	9.5
Tetrachloro-m-xylene	3.660	3.560	3.760	52.490	50.000	5.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q3100
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/15/2025 09/15/2025

Client Sample No.: CCAL03 **Date Analyzed:** 09/16/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113784.D **Time Analyzed:** 23:32

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.722	4.623	4.823	475.680	500.000	-4.9
Aroclor-1016-2	4.740	4.641	4.841	481.080	500.000	-3.8
Aroclor-1016-3	4.915	4.816	5.016	492.920	500.000	-1.4
Aroclor-1016-4	4.957	4.858	5.058	486.620	500.000	-2.7
Aroclor-1016-5	5.169	5.070	5.270	511.810	500.000	2.4
Aroclor-1260-1	6.194	6.096	6.296	473.580	500.000	-5.3
Aroclor-1260-2	6.383	6.284	6.484	473.370	500.000	-5.3
Aroclor-1260-3	6.534	6.435	6.635	450.050	500.000	-10.0
Aroclor-1260-4	7.002	6.903	7.103	470.560	500.000	-5.9
Aroclor-1260-5	7.245	7.146	7.346	465.330	500.000	-6.9
Decachlorobiphenyl	8.614	8.514	8.714	51.340	50.000	2.7
Tetrachloro-m-xylene	3.651	3.551	3.751	50.000	50.000	0.0