

Surrogate Summary

SDG No.: **Q3100**

Client: **ATC Group Services LLC**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PO113719.D	PIBLK-PO113719.D	Tetrachloro-m-xyl	1	20	17.0	85		60	140
		Decachlorobiphen	1	20	17.3	86		60	140
		Tetrachloro-m-xyl	2	20	17.0	85		60	140
		Decachlorobiphen	2	20	18.1	91		60	140
I.BLK-PO113759.D	PIBLK-PO113759.D	Tetrachloro-m-xyl	1	20	17.5	87		60	140
		Decachlorobiphen	1	20	20.5	102		60	140
		Tetrachloro-m-xyl	2	20	17.3	87		60	140
		Decachlorobiphen	2	20	19.9	100		60	140
PB169705BL	PB169705BL	Tetrachloro-m-xyl	1	20	18.6	93		32	144
		Decachlorobiphen	1	20	21.8	109		32	175
		Tetrachloro-m-xyl	2	20	17.9	90		32	144
		Decachlorobiphen	2	20	20.5	102		32	175
PB169705BS	PB169705BS	Tetrachloro-m-xyl	1	20	17.7	89		32	144
		Decachlorobiphen	1	20	21.7	109		32	175
		Tetrachloro-m-xyl	2	20	17.0	85		32	144
		Decachlorobiphen	2	20	20.5	103		32	175
I.BLK-PO113774.D	PIBLK-PO113774.D	Tetrachloro-m-xyl	1	20	18.1	91		60	140
		Decachlorobiphen	1	20	19.1	96		60	140
		Tetrachloro-m-xyl	2	20	17.5	87		60	140
		Decachlorobiphen	2	20	18.6	93		60	140
Q3100-01	ROOF	Tetrachloro-m-xyl	1	20	18.5	93		32	144
		Decachlorobiphen	1	20	20.6	103		32	175
		Tetrachloro-m-xyl	2	20	17.7	88		32	144
		Decachlorobiphen	2	20	17.7	89		32	175
Q3100-02	2A2B2C	Tetrachloro-m-xyl	1	20	18.6	93		32	144
		Decachlorobiphen	1	20	21.5	108		32	175
		Tetrachloro-m-xyl	2	20	19.0	95		32	144
		Decachlorobiphen	2	20	20.9	104		32	175
Q3100-03	3A3B3C	Tetrachloro-m-xyl	1	20	17.0	85		32	144
		Decachlorobiphen	1	20	21.2	106		32	175
		Tetrachloro-m-xyl	2	20	17.8	89		32	144
		Decachlorobiphen	2	20	20.7	104		32	175
I.BLK-PO113788.D	PIBLK-PO113788.D	Tetrachloro-m-xyl	1	20	18.6	93		60	140
		Decachlorobiphen	1	20	21.8	109		60	140
		Tetrachloro-m-xyl	2	20	17.9	90		60	140
		Decachlorobiphen	2	20	19.6	98		60	140