


LB137203

7471B

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 1.9440e-004

C= 7.1890e-002

Rho= 0.9998298

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%CV
0.0	0.000	0.082	0.082	53	0.000	53					
0.2	0.200	0.231	0.031	819	0.0 %	819					
2.5	2.500	2.375	-0.125	11845	0.0 %	11845					
5.0	5.000	4.982	-0.018	25257	0.0 %	25257					
7.5	7.500	7.473	-0.027	38071	0.0 %	38071					
10.0	10.000	10.057	0.057	51365	0.0 %	51365					

0/01)
16
-5
00
1

LB137203 INSTRUMENT ID : CV1

Method: 7471B Operator: Admin Date of Analysis: 16 Sep 2025 13:35:52

Type	Sample ID	Extended ID	Conc.	μ Abs.	Units	Type	Std Conc	Method	Date
S	0.0 - 1	50	-	53	PPB	Std	0.0000	7471B	16 Sep 2025 13:38:05
S	0.2 - 1	50.2	-	819	PPB	Std	0.2000	7471B	16 Sep 2025 13:40:22
S	2.5 - 1	52.5	-	11845	PPB	Std	2.5000	7471B	16 Sep 2025 13:42:39
S	5.0 - 1	55	-	25257	PPB	Std	5.0000	7471B	16 Sep 2025 13:44:56
S	7.5 - 1	57.5	-	38071	PPB	Std	7.5000	7471B	16 Sep 2025 13:47:26
S	10.0 - 1	510	-	51365	PPB	Std	10.0000	7471B	16 Sep 2025 13:50:01
U	ICV31 - 1	ICV31	4.0093	20254	PPB	SMPL	-7471B		16 Sep 2025 13:55:55
U	ICB31 - 1	ICB31	0.0219	-257	PPB	SMPL	-7471B		16 Sep 2025 13:58:10
U	CCV95 - 1	CCV95	4.9605	25147	PPB	SMPL	-7471B		16 Sep 2025 14:00:44
U	CCB95 - 1	CCB95	0.0315	-208	PPB	SMPL	-7471B		16 Sep 2025 14:02:59
U	CRA - 1	CRA	0.2268	797	PPB	SMPL	-7471B		16 Sep 2025 14:05:36
U	HighStd - 1	HighStd	9.9065	50589	PPB	SMPL	-7471B		16 Sep 2025 14:07:51
U	ChkStd - 1	ChkStd	6.7025	34108	PPB	SMPL	-7471B		16 Sep 2025 14:10:06
U	PB169707BL - 1	PBS	0.0283	-224	PPB	SMPL	-7471B		16 Sep 2025 14:12:53
U	PB169707BS - 1	LCSS	3.9685	20044	PPB	SMPL	-7471B		16 Sep 2025 14:20:53
U	Q3101-01 - 1	Lot#S:091/15	0.0453	-137	PPB	SMPL	-7471B		16 Sep 2025 14:23:15
U	Q3104-01 - 1	SU-04-091525	1.2680	6153	PPB	SMPL	-7471B		16 Sep 2025 14:25:51
U	Q3104-01DUP - 1	SU-04-091525DUP	1.0878	5226	PPB	SMPL	-7471B		16 Sep 2025 14:28:21
U	Q3104-01MS - 1	SU-04-091525MS	4.9654	25172	PPB	SMPL	-7471B		16 Sep 2025 14:33:50
U	Q3104-01MSD - 1	SU-04-091525MSD	4.8600	24630	PPB	SMPL	-7471B		16 Sep 2025 14:36:08
U	CCV96 - 1	CCV96	4.8958	24814	PPB	SMPL	-7471B		16 Sep 2025 14:38:43
U	CCB96 - 1	CCB96	0.0559	-82	PPB	SMPL	-7471B		16 Sep 2025 14:41:21
U	Q3105-01 - 1	FREEZE-PIT-SAMPLES	0.1613	460	PPB	SMPL	-7471B		16 Sep 2025 14:44:12
U	Q3106-01 - 1	NB-07-09152025	0.3509	1435	PPB	SMPL	-7471B		16 Sep 2025 14:46:35
U	Q3106-03 - 1	NB-08-09152025	0.1568	437	PPB	SMPL	-7471B		16 Sep 2025 14:48:50
U	Q3104-01LX5 - 1		0.2546	940	PPB	SMPL	-7471B		16 Sep 2025 14:57:28
U	Q3104-01A - 1		4.7521	24075	PPB	SMPL	-7471B		16 Sep 2025 14:59:50
U	CCV97 - 1	CCV97	4.9309	24995	PPB	SMPL	-7471B		16 Sep 2025 15:02:14
U	CCB97 - 1	CCB97	0.0313	-209	PPB	SMPL	-7471B		16 Sep 2025 15:04:51