



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Vermont's Original, LLC

SDG No.: Q3101

Contract: DAIR01

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Aluminum	111	100	110	65 - 135	P	09/16/2025	11:38	LB137209
	Antimony	48.7	50.0	98	65 - 135	P	09/16/2025	11:38	LB137209
	Arsenic	19.9	20.0	100	65 - 135	P	09/16/2025	11:38	LB137209
	Barium	92.8	100	93	65 - 135	P	09/16/2025	11:38	LB137209
	Beryllium	5.93	6.0	99	65 - 135	P	09/16/2025	11:38	LB137209
	Cadmium	5.49	6.0	92	65 - 135	P	09/16/2025	11:38	LB137209
	Calcium	1990	2000	100	65 - 135	P	09/16/2025	11:38	LB137209
	Chromium	10.00	10.0	100	65 - 135	P	09/16/2025	11:38	LB137209
	Cobalt	28.3	30.0	94	65 - 135	P	09/16/2025	11:38	LB137209
	Copper	21.0	20.0	105	65 - 135	P	09/16/2025	11:38	LB137209
	Iron	98.2	100	98	65 - 135	P	09/16/2025	11:38	LB137209
	Lead	11.4	12.0	95	65 - 135	P	09/16/2025	11:38	LB137209
	Magnesium	2020	2000	101	65 - 135	P	09/16/2025	11:38	LB137209
	Manganese	20.5	20.0	102	65 - 135	P	09/16/2025	11:38	LB137209
	Nickel	38.3	40.0	96	65 - 135	P	09/16/2025	11:38	LB137209
	Potassium	1860	2000	93	65 - 135	P	09/16/2025	11:38	LB137209
	Selenium	22.6	20.0	113	65 - 135	P	09/16/2025	11:38	LB137209
	Silver	10.4	10.0	104	65 - 135	P	09/16/2025	11:38	LB137209
	Sodium	1760	2000	88	65 - 135	P	09/16/2025	11:38	LB137209
	Thallium	35.6	40.0	89	65 - 135	P	09/16/2025	11:38	LB137209
	Vanadium	38.5	40.0	96	65 - 135	P	09/16/2025	11:38	LB137209
	Zinc	43.1	40.0	108	65 - 135	P	09/16/2025	11:38	LB137209
CRA	Mercury	0.23	0.2	113	70 - 130	CV	09/16/2025	14:05	LB137203