



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Vermont's Original, LLC

SDG No.: Q3101

Contract: DAIR01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB31	Mercury	0.076	+/-0.2	U	0.20	CV	09/16/2025	13:58	LB137203

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Vermont's Original, LLC

SDG No.: Q3101

Contract: DAIR01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB95	Mercury	0.076	+/-0.2	U	0.20	CV	09/16/2025	14:02	LB137203
CCB96	Mercury	0.076	+/-0.2	U	0.20	CV	09/16/2025	14:41	LB137203
CCB97	Mercury	0.076	+/-0.2	U	0.20	CV	09/16/2025	15:04	LB137203

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Vermont's Original, LLC

SDG No.: Q3101

Contract: DAIR01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	14.5	+/-50	J	100	P	09/16/2025	11:33	LB137209
	Antimony	6.76	+/-25	U	50.0	P	09/16/2025	11:33	LB137209
	Arsenic	5.12	+/-10	U	20.0	P	09/16/2025	11:33	LB137209
	Barium	14.6	+/-50	U	100	P	09/16/2025	11:33	LB137209
	Beryllium	0.56	+/-3	U	6.00	P	09/16/2025	11:33	LB137209
	Cadmium	0.50	+/-3	U	6.00	P	09/16/2025	11:33	LB137209
	Calcium	234	+/-1000	U	2000	P	09/16/2025	11:33	LB137209
	Chromium	2.12	+/-5	U	10.0	P	09/16/2025	11:33	LB137209
	Cobalt	2.26	+/-15	U	30.0	P	09/16/2025	11:33	LB137209
	Copper	4.60	+/-10	U	20.0	P	09/16/2025	11:33	LB137209
	Iron	23.4	+/-50	U	100	P	09/16/2025	11:33	LB137209
	Lead	2.30	+/-6	U	12.0	P	09/16/2025	11:33	LB137209
	Magnesium	244	+/-1000	U	2000	P	09/16/2025	11:33	LB137209
	Manganese	5.94	+/-10	U	20.0	P	09/16/2025	11:33	LB137209
	Nickel	3.06	+/-20	U	40.0	P	09/16/2025	11:33	LB137209
	Potassium	918	+/-1000	U	2000	P	09/16/2025	11:33	LB137209
	Selenium	9.64	+/-10	U	20.0	P	09/16/2025	11:33	LB137209
	Silver	1.62	+/-5	U	10.0	P	09/16/2025	11:33	LB137209
	Sodium	868	+/-1000	U	2000	P	09/16/2025	11:33	LB137209
	Thallium	4.38	+/-20	U	40.0	P	09/16/2025	11:33	LB137209
	Vanadium	6.26	+/-20	U	40.0	P	09/16/2025	11:33	LB137209
	Zinc	16.7	+/-20	U	40.0	P	09/16/2025	11:33	LB137209

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Vermont's Original, LLC

SDG No.: Q3101

Contract: DAIR01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	13.9	+/-50	J	100	P	09/16/2025	12:05	LB137209
	Antimony	6.76	+/-25	U	50.0	P	09/16/2025	12:05	LB137209
	Arsenic	5.12	+/-10	U	20.0	P	09/16/2025	12:05	LB137209
	Barium	14.6	+/-50	U	100	P	09/16/2025	12:05	LB137209
	Beryllium	0.56	+/-3	U	6.00	P	09/16/2025	12:05	LB137209
	Cadmium	0.50	+/-3	U	6.00	P	09/16/2025	12:05	LB137209
	Calcium	234	+/-1000	U	2000	P	09/16/2025	12:05	LB137209
	Chromium	2.12	+/-5	U	10.0	P	09/16/2025	12:05	LB137209
	Cobalt	2.26	+/-15	U	30.0	P	09/16/2025	12:05	LB137209
	Copper	4.60	+/-10	U	20.0	P	09/16/2025	12:05	LB137209
	Iron	23.4	+/-50	U	100	P	09/16/2025	12:05	LB137209
	Lead	2.30	+/-6	U	12.0	P	09/16/2025	12:05	LB137209
	Magnesium	244	+/-1000	U	2000	P	09/16/2025	12:05	LB137209
	Manganese	5.94	+/-10	U	20.0	P	09/16/2025	12:05	LB137209
	Nickel	3.06	+/-20	U	40.0	P	09/16/2025	12:05	LB137209
	Potassium	918	+/-1000	U	2000	P	09/16/2025	12:05	LB137209
	Selenium	9.64	+/-10	U	20.0	P	09/16/2025	12:05	LB137209
	Silver	1.62	+/-5	U	10.0	P	09/16/2025	12:05	LB137209
	Sodium	868	+/-1000	U	2000	P	09/16/2025	12:05	LB137209
	Thallium	4.38	+/-20	U	40.0	P	09/16/2025	12:05	LB137209
	Vanadium	6.26	+/-20	U	40.0	P	09/16/2025	12:05	LB137209
	Zinc	16.7	+/-20	U	40.0	P	09/16/2025	12:05	LB137209
CCB02	Aluminum	16.9	+/-50	J	100	P	09/16/2025	13:43	LB137209
	Antimony	6.76	+/-25	U	50.0	P	09/16/2025	13:43	LB137209
	Arsenic	5.12	+/-10	U	20.0	P	09/16/2025	13:43	LB137209
	Barium	14.6	+/-50	U	100	P	09/16/2025	13:43	LB137209
	Beryllium	0.56	+/-3	U	6.00	P	09/16/2025	13:43	LB137209
	Cadmium	0.50	+/-3	U	6.00	P	09/16/2025	13:43	LB137209
	Calcium	234	+/-1000	U	2000	P	09/16/2025	13:43	LB137209
	Chromium	2.12	+/-5	U	10.0	P	09/16/2025	13:43	LB137209
	Cobalt	2.26	+/-15	U	30.0	P	09/16/2025	13:43	LB137209
	Copper	4.60	+/-10	U	20.0	P	09/16/2025	13:43	LB137209
	Iron	23.4	+/-50	U	100	P	09/16/2025	13:43	LB137209
	Lead	2.30	+/-6	U	12.0	P	09/16/2025	13:43	LB137209
	Magnesium	244	+/-1000	U	2000	P	09/16/2025	13:43	LB137209
	Manganese	5.94	+/-10	U	20.0	P	09/16/2025	13:43	LB137209
	Nickel	3.06	+/-20	U	40.0	P	09/16/2025	13:43	LB137209
	Potassium	918	+/-1000	U	2000	P	09/16/2025	13:43	LB137209
	Selenium	9.64	+/-10	U	20.0	P	09/16/2025	13:43	LB137209

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Vermont's Original, LLC

SDG No.: Q3101

Contract: DAIR01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	1.62	+/-5	U	10.0	P	09/16/2025	13:43	LB137209
	Sodium	868	+/-1000	U	2000	P	09/16/2025	13:43	LB137209
	Thallium	4.38	+/-20	U	40.0	P	09/16/2025	13:43	LB137209
	Vanadium	6.26	+/-20	U	40.0	P	09/16/2025	13:43	LB137209
	Zinc	16.7	+/-20	U	40.0	P	09/16/2025	13:43	LB137209
CCB03	Aluminum	16.1	+/-50	J	100	P	09/16/2025	14:47	LB137209
	Antimony	6.76	+/-25	U	50.0	P	09/16/2025	14:47	LB137209
	Arsenic	5.12	+/-10	U	20.0	P	09/16/2025	14:47	LB137209
	Barium	14.6	+/-50	U	100	P	09/16/2025	14:47	LB137209
	Beryllium	0.56	+/-3	U	6.00	P	09/16/2025	14:47	LB137209
	Cadmium	0.50	+/-3	U	6.00	P	09/16/2025	14:47	LB137209
	Calcium	234	+/-1000	U	2000	P	09/16/2025	14:47	LB137209
	Chromium	2.12	+/-5	U	10.0	P	09/16/2025	14:47	LB137209
	Cobalt	2.26	+/-15	U	30.0	P	09/16/2025	14:47	LB137209
	Copper	4.60	+/-10	U	20.0	P	09/16/2025	14:47	LB137209
	Iron	23.4	+/-50	U	100	P	09/16/2025	14:47	LB137209
	Lead	2.30	+/-6	U	12.0	P	09/16/2025	14:47	LB137209
	Magnesium	244	+/-1000	U	2000	P	09/16/2025	14:47	LB137209
	Manganese	5.94	+/-10	U	20.0	P	09/16/2025	14:47	LB137209
	Nickel	3.06	+/-20	U	40.0	P	09/16/2025	14:47	LB137209
	Potassium	918	+/-1000	U	2000	P	09/16/2025	14:47	LB137209
	Selenium	9.64	+/-10	U	20.0	P	09/16/2025	14:47	LB137209
	Silver	1.62	+/-5	U	10.0	P	09/16/2025	14:47	LB137209
	Sodium	868	+/-1000	U	2000	P	09/16/2025	14:47	LB137209
	Thallium	4.38	+/-20	U	40.0	P	09/16/2025	14:47	LB137209
	Vanadium	6.26	+/-20	U	40.0	P	09/16/2025	14:47	LB137209
	Zinc	16.7	+/-20	U	40.0	P	09/16/2025	14:47	LB137209
CCB04	Aluminum	12.8	+/-50	J	100	P	09/16/2025	15:36	LB137209
	Antimony	6.76	+/-25	U	50.0	P	09/16/2025	15:36	LB137209
	Arsenic	5.12	+/-10	U	20.0	P	09/16/2025	15:36	LB137209
	Barium	14.6	+/-50	U	100	P	09/16/2025	15:36	LB137209
	Beryllium	0.56	+/-3	U	6.00	P	09/16/2025	15:36	LB137209
	Cadmium	0.50	+/-3	U	6.00	P	09/16/2025	15:36	LB137209
	Calcium	234	+/-1000	U	2000	P	09/16/2025	15:36	LB137209
	Chromium	2.12	+/-5	U	10.0	P	09/16/2025	15:36	LB137209
	Cobalt	2.26	+/-15	U	30.0	P	09/16/2025	15:36	LB137209
	Copper	4.60	+/-10	U	20.0	P	09/16/2025	15:36	LB137209
	Iron	23.4	+/-50	U	100	P	09/16/2025	15:36	LB137209
	Lead	2.30	+/-6	U	12.0	P	09/16/2025	15:36	LB137209

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Vermont's Original, LLC

SDG No.: Q3101

Contract: DAIR01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	244	+/-1000	U	2000	P	09/16/2025	15:36	LB137209
	Manganese	5.94	+/-10	U	20.0	P	09/16/2025	15:36	LB137209
	Nickel	3.06	+/-20	U	40.0	P	09/16/2025	15:36	LB137209
	Potassium	918	+/-1000	U	2000	P	09/16/2025	15:36	LB137209
	Selenium	9.64	+/-10	U	20.0	P	09/16/2025	15:36	LB137209
	Silver	1.62	+/-5	U	10.0	P	09/16/2025	15:36	LB137209
	Sodium	868	+/-1000	U	2000	P	09/16/2025	15:36	LB137209
	Thallium	4.38	+/-20	U	40.0	P	09/16/2025	15:36	LB137209
	Vanadium	6.26	+/-20	U	40.0	P	09/16/2025	15:36	LB137209
	Zinc	16.7	+/-20	U	40.0	P	09/16/2025	15:36	LB137209
CCB05	Aluminum	26.4	+/-50	J	100	P	09/16/2025	16:31	LB137209
	Antimony	6.76	+/-25	U	50.0	P	09/16/2025	16:31	LB137209
	Arsenic	5.12	+/-10	U	20.0	P	09/16/2025	16:31	LB137209
	Barium	14.6	+/-50	U	100	P	09/16/2025	16:31	LB137209
	Beryllium	0.56	+/-3	U	6.00	P	09/16/2025	16:31	LB137209
	Cadmium	0.50	+/-3	U	6.00	P	09/16/2025	16:31	LB137209
	Calcium	234	+/-1000	U	2000	P	09/16/2025	16:31	LB137209
	Chromium	2.12	+/-5	U	10.0	P	09/16/2025	16:31	LB137209
	Cobalt	2.26	+/-15	U	30.0	P	09/16/2025	16:31	LB137209
	Copper	4.60	+/-10	U	20.0	P	09/16/2025	16:31	LB137209
	Iron	23.4	+/-50	U	100	P	09/16/2025	16:31	LB137209
	Lead	2.30	+/-6	U	12.0	P	09/16/2025	16:31	LB137209
	Magnesium	244	+/-1000	U	2000	P	09/16/2025	16:31	LB137209
	Manganese	5.94	+/-10	U	20.0	P	09/16/2025	16:31	LB137209
	Nickel	3.06	+/-20	U	40.0	P	09/16/2025	16:31	LB137209
	Potassium	918	+/-1000	U	2000	P	09/16/2025	16:31	LB137209
	Selenium	9.64	+/-10	U	20.0	P	09/16/2025	16:31	LB137209
	Silver	1.62	+/-5	U	10.0	P	09/16/2025	16:31	LB137209
	Sodium	868	+/-1000	U	2000	P	09/16/2025	16:31	LB137209
	Thallium	4.38	+/-20	U	40.0	P	09/16/2025	16:31	LB137209
	Vanadium	6.26	+/-20	U	40.0	P	09/16/2025	16:31	LB137209
	Zinc	16.7	+/-20	U	40.0	P	09/16/2025	16:31	LB137209
CCB06	Aluminum	13.0	+/-50	J	100	P	09/16/2025	16:53	LB137209
	Antimony	6.76	+/-25	U	50.0	P	09/16/2025	16:53	LB137209
	Arsenic	5.12	+/-10	U	20.0	P	09/16/2025	16:53	LB137209
	Barium	14.6	+/-50	U	100	P	09/16/2025	16:53	LB137209
	Beryllium	0.56	+/-3	U	6.00	P	09/16/2025	16:53	LB137209
	Cadmium	0.50	+/-3	U	6.00	P	09/16/2025	16:53	LB137209
	Calcium	234	+/-1000	U	2000	P	09/16/2025	16:53	LB137209

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Vermont's Original, LLC

SDG No.: Q3101

Contract: DAIR01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	2.12	+/-5	U	10.0	P	09/16/2025	16:53	LB137209
	Cobalt	2.26	+/-15	U	30.0	P	09/16/2025	16:53	LB137209
	Copper	4.60	+/-10	U	20.0	P	09/16/2025	16:53	LB137209
	Iron	23.4	+/-50	U	100	P	09/16/2025	16:53	LB137209
	Lead	2.30	+/-6	U	12.0	P	09/16/2025	16:53	LB137209
	Magnesium	244	+/-1000	U	2000	P	09/16/2025	16:53	LB137209
	Manganese	5.94	+/-10	U	20.0	P	09/16/2025	16:53	LB137209
	Nickel	3.06	+/-20	U	40.0	P	09/16/2025	16:53	LB137209
	Potassium	918	+/-1000	U	2000	P	09/16/2025	16:53	LB137209
	Selenium	9.64	+/-10	U	20.0	P	09/16/2025	16:53	LB137209
	Silver	1.62	+/-5	U	10.0	P	09/16/2025	16:53	LB137209
	Sodium	868	+/-1000	U	2000	P	09/16/2025	16:53	LB137209
	Thallium	4.38	+/-20	U	40.0	P	09/16/2025	16:53	LB137209
	Vanadium	6.26	+/-20	U	40.0	P	09/16/2025	16:53	LB137209
	Zinc	16.7	+/-20	U	40.0	P	09/16/2025	16:53	LB137209

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Vermont's Original, LLC

SDG No.: Q3101

Contract: DAIR01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
-----------	---------	----------------	---------------------	--------------	------	---	------------------	------------------	---------------