



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

### Initial and Continuing Calibration Verification

**Client:** Urban Auto Spa Car Wash

**SDG No.:** Q3102

**Project:** Annual 2024 - UAS of Wayne, NJ

**RunNo.:** LB137258

| Analyte    |      | Units | Result | True Value | %<br>Recovery | Acceptance<br>Window (%R) | Analysis<br>Date |
|------------|------|-------|--------|------------|---------------|---------------------------|------------------|
| Sample ID: | ICV  |       |        |            |               |                           |                  |
| Field pH   |      | pH    | 6.99   | 7          | 100           | 90-110                    | 09/15/2025       |
| Sample ID: | CCV1 |       |        |            |               |                           |                  |
| Field pH   |      | pH    | 6.99   | 7          | 100           | 90-110                    | 09/15/2025       |
| Sample ID: | CCV2 |       |        |            |               |                           |                  |
| Field pH   |      | pH    | 7.00   | 7          | 100           | 90-110                    | 09/15/2025       |

## Initial and Continuing Calibration Verification

**Client:** Urban Auto Spa Car Wash

**SDG No.:** Q3102

**Project:** Annual 2024 - UAS of Wayne, NJ

**RunNo.:** LB137264

| Analyte                       | Units | Result | True Value | % Recovery | Acceptance Window (%R) | Analysis Date |
|-------------------------------|-------|--------|------------|------------|------------------------|---------------|
| Sample ID: <b>ICV</b><br>COD  | mg/L  | 50.173 | 50         | 100        | 95-105                 | 09/08/2025    |
| Sample ID: <b>CCV1</b><br>COD | mg/L  | 51.184 | 50         | 102        | 95-105                 | 09/22/2025    |
| Sample ID: <b>CCV2</b><br>COD | mg/L  | 50.173 | 50         | 100        | 95-105                 | 09/22/2025    |