

Duplicate Sample Summary

Client:	Urban Auto Spa Car Wash	SDG No.:	Q3102
Project:	Annual 2024 - UAS of Wayne, NJ	Sample ID:	LB137253BS
Client ID:	LB137253BSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TPH	mg/L	+/-18	16.6		17.0		1	2.38		09/22/2025

Duplicate Sample Summary

Client:	Urban Auto Spa Car Wash	SDG No.:	Q3102
Project:	Annual 2024 - UAS of Wayne, NJ	Sample ID:	Q3102-01
Client ID:	UAS-WAYNEDUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
CBOD5	mg/L	+/-20	116		121		1	3.8		09/17/2025
TSS	mg/L	+/-5	93.2		93.1		1	0.11		09/19/2025
Field pH	pH	+/-20	6.97		6.99		1	0.29		09/15/2025

Duplicate Sample Summary

Client:	Urban Auto Spa Car Wash	SDG No.:	Q3102
Project:	Annual 2024 - UAS of Wayne, NJ	Sample ID:	Q3102-01
Client ID:	UAS-WAYNEMSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Oil and Grease	mg/L	+/-18	58.4		58.5		1	0.17		09/22/2025

Duplicate Sample Summary

Client:	Urban Auto Spa Car Wash	SDG No.:	Q3102
Project:	Annual 2024 - UAS of Wayne, NJ	Sample ID:	Q3151-02
Client ID:	1402DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
COD	mg/L	+/-20	2.64	J	2.64	J	1	0		09/22/2025

Duplicate Sample Summary

Client:	Urban Auto Spa Car Wash	SDG No.:	Q3102
Project:	Annual 2024 - UAS of Wayne, NJ	Sample ID:	Q3151-02
Client ID:	1402MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
COD	mg/L	+/-20	48.2		49.2		1	2.05		09/22/2025