

CLIENT INFORMATION

CLIENT PROJECT INFORMATION

CLIENT BILLING INFORMATION

COMPANY: Urban Auto Spa
ADDRESS: 755 Hamburg Turnpike
CITY: Wayne STATE: NJ ZIP:
ATTENTION: Edward Dunn
PHONE: FAX:

PROJECT NAME: Urban Auto Spa - Wayne
PROJECT NO.: LOCATION:
PROJECT MANAGER: Edward Dunn
e-mail:
PHONE: FAX:

BILL TO: PO#:
ADDRESS:
CITY: STATE: ZIP:
ATTENTION: PHONE:

ANALYSIS

DATA TURNAROUND INFORMATION

DATA DELIVERABLE INFORMATION

FAX (RUSH) DAYS*
HARDCOPY (DATA PACKAGE): DAYS*
EDD: DAYS*

*TO BE APPROVED BY CHEMTECH

STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)
☐ Level 2 (Results + QC) ☐ NJ Reduced ☐ US EPA CLP
☐ Level 3 (Results + QC) ☐ NYS ASP A ☐ NYS ASP B
+ Raw Data ☐ Other
☐ EDD FORMAT

Oil And Grease
TPH
TSS
CBOD5
COD
1 2 3 4 5 6 7 8 9

PRESERVATIVES

COMMENTS

ALLIANCE SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# OF BOTTLES										COMMENTS
			COMP	GRAB	DATE	TIME		1	2	3	4	5	6	7	8	9	
1.	UAS-Wayne	AQ		X	9.15.25	946	8	X	X	X	X	X					PH=6.97
2.																	
3.																	
4.																	
5.																	
6.																	
7.																	
8.																	
9.																	
10.																	

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER: 1. <u>JT</u>	DATE/TIME: <u>1020</u> <u>9.15.25</u>	RECEIVED BY: 1. <u></u>	Conditions of bottles or coolers at receipt: ☒ COMPLIANT ☐ NON COMPLIANT ☐ COOLER TEMP <u>3.4</u> °C
RELINQUISHED BY SAMPLER: 2. <u></u>	DATE/TIME: <u></u>	RECEIVED BY: 2. <u></u>	Comments: <u></u>
RELINQUISHED BY SAMPLER: 3. <u>JT</u>	DATE/TIME: <u>1300</u> <u>9.15.25</u>	RECEIVED BY: 3. <u></u>	Page <u></u> of <u></u> CLIENT: ☐ Hand Delivered ☐ Other <u></u> Shipment Complete ☐ YES ☐ NO

Project Name: UAS of Wayne, NJ
Project Location: Wayne, NJ
Cooler Custody Seal: NA
Temperature Correction Factor (°C): 0

[illegible]

Sampler Signature/Date:

Supervisor Review/Date:

FIELD SAMPLING LOGClient Name: Wayne - URBAN Auto SPAClient Address: 755 Hamburg TurnpikeClient Rep on Site: Edward DunnSampling Date: 9.15-25Arrival Time: 920Departure Time: 1020Project Name: UAS of Wayne, NJProject Location: Wayne, NJCooler Custody Seal: N/ATemperature Correction Factor (°C): 0FIELD EQUIPMENT CALIBRATION

pH Calibration (SM4500-H B/9040C)				
Calibration				ICV (± 0.1 pH unit)
	7.00 Buffer	4.00 Buffer	10.00 Buffer	7.00 Buffer
	W 3217	W 3178	W 3191	W 3093
Time	931	933	935	937
Temp °C	21.2	21.2	21.3	21.1
pH	6.96	3.99	9.95	6.99

FIELD EQUIPMENT CALIBRATION

Specific Conductance (mS/cm) (99% -101%)/(mmho/cm) (SM2510 B/120.1/9050A)		
Calibration (± 1%) (99% -101%)		ICV (± 1%) (99% -101%)
	WP	WP
Time		
Temp °C		
Reading (mS/cm)		

Sampler Signature/Date: J. Bragg 9.15-25Supervisor Review/Date: Se 9/15/25