

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB137222

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3194

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.7	09/18/2025	09:40
2	Q3103-01		1	15	100.00	>212.0	>212.0	09/18/2025	10:10
3	Q3103-01DUP		1	15	100.00	>212.0	>212.0	09/18/2025	11:00
4	Q3113-01		1	13	100.00	>212.0	>212.0	09/18/2025	11:30
5	Q3116-01		1	15	100.00	>212.0	>212.0	09/18/2025	12:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06