

Duplicate Sample Summary

Client: ENTACT Project: 540 Degraw St, Brooklyn, NY - E9309 Client ID: UAS-WAYNEDUP	SDG No.: Q3103 Sample ID: Q3102-01 Percent Solids for Spike Sample: 0
--	--

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TSS	mg/L	+/-5	93.2		93.1		1	0.11		09/19/2025

Duplicate Sample Summary

Client:	ENTACT	SDG No.:	Q3103
Project:	540 Degraw St, Brooklyn, NY - E9309	Sample ID:	Q3103-01
Client ID:	TW-WTS-14DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
BOD5	mg/L	+/-20	18.6		19.6		1	4.87		09/17/2025
Flash Point	o F	+/-2	>212.0		>212.0		1	0		09/18/2025