

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: Q3103

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB137215

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137215BS							
BOD5	mg/L	198	173		87	1	84.6-115.4	09/17/2025

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: Q3103

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB137250

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137250BS							
TSS	mg/L	550	532		97	1	90-110	09/19/2025