



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: ENTACT

SDG No.: Q3103

Contract: ENTA05

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.19	0.2	96	70 - 130	CV	09/17/2025	14:30	LB137213
CRI01	Cadmium	5.68	6.0	95	65 - 135	P	09/18/2025	11:28	LB137227
	Copper	21.3	20.0	106	65 - 135	P	09/18/2025	11:28	LB137227
	Lead	11.6	12.0	97	65 - 135	P	09/18/2025	11:28	LB137227
	Nickel	39.8	40.0	99	65 - 135	P	09/18/2025	11:28	LB137227
	Zinc	44.6	40.0	111	65 - 135	P	09/18/2025	11:28	LB137227