

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: ENTACT

SDG No.: Q3103

Contract: ENTA05

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Cadmium	-0.069	1.0	7	-5	7	09/18/2025	11:33	LB137227
	Copper	19.7	2.0	985	-18	22	09/18/2025	11:33	LB137227
	Lead	0.77			-12	12	09/18/2025	11:33	LB137227
	Nickel	7.93	2.0	396	-38	42	09/18/2025	11:33	LB137227
	Zinc	3.87			-40	40	09/18/2025	11:33	LB137227
ICSAB01	Cadmium	1010	972	104	826	1120	09/18/2025	11:38	LB137227
	Copper	515	511	101	434	588	09/18/2025	11:38	LB137227
	Lead	49.1	49.0	100	37	61	09/18/2025	11:38	LB137227
	Nickel	1020	954	107	810	1100	09/18/2025	11:38	LB137227
	Zinc	1060	952	111	809	1095	09/18/2025	11:38	LB137227