

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: ENTACT

SDG No.: Q3103

Contract: ENTA05

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169721BS							
Cadmium	ug/L	100	91.7		92	80 - 120	P
Copper	ug/L	150	148		99	80 - 120	P
Lead	ug/L	500	458		92	80 - 120	P
Nickel	ug/L	250	234		94	80 - 120	P
Zinc	ug/L	100	97.5		98	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: ENTACT

SDG No.: Q3103

Contract: ENTA05

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169732BS Mercury	ug/L	4.0	4.33		108	80 - 120	CV