

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ENTA05

Lab Code: ACE

SDG NO.: Q3103

Continuing Calib Date: 09/16/2025

Initial Calibration Date(s): 09/15/2025

09/15/2025

Continuing Calib Time: 11:04

Initial Calibration Time(s): 09:34

19:21

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1254-1 (1)	5.55	5.55	5.45	5.65	0.00
Aroclor-1254-2 (2)	5.70	5.70	5.60	5.80	0.00
Aroclor-1254-3 (3)	6.10	6.10	6.00	6.20	0.00
Aroclor-1254-4 (4)	6.33	6.33	6.23	6.43	0.00
Aroclor-1254-5 (5)	6.74	6.75	6.65	6.85	0.01
Tetrachloro-m-xylene	3.66	3.66	3.56	3.76	0.00
Decachlorobiphenyl	8.67	8.67	8.57	8.77	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ENTA05

Lab Code: ACE

SDG NO.: Q3103

Continuing Calib Date: 09/16/2025

Initial Calibration Date(s): 09/15/2025 09/15/2025

Continuing Calib Time: 11:04

Initial Calibration Time(s): 09:34 19:21

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1254-1 (1)	5.52	5.52	5.42	5.62	0.00
Aroclor-1254-2 (2)	5.67	5.67	5.57	5.77	0.01
Aroclor-1254-3 (3)	6.07	6.07	5.97	6.17	0.00
Aroclor-1254-4 (4)	6.29	6.29	6.19	6.39	0.00
Aroclor-1254-5 (5)	6.71	6.71	6.61	6.81	0.00
Tetrachloro-m-xylene	3.65	3.65	3.55	3.75	0.00
Decachlorobiphenyl	8.61	8.62	8.52	8.72	0.01