

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: ENTA05
 Lab Code: ACE SDG No.: Q3103
 Instrument ID: BNA_P Calibration Date/Time: 09/17/2025 12:03
 Lab File ID: BP025769.D Init. Calib. Date(s): 09/11/2025 09/11/2025
 EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 08:56 15:20
 GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.240	1.323		6.7	
Phenol-d6	1.648	1.712		3.9	
Phenol	1.737	1.782		2.6	20.0
1,4-Dichlorobenzene	1.576	1.623		3.0	20.0
Nitrobenzene-d5	0.453	0.476		5.1	
1,2,4-Trichlorobenzene	0.356	0.383		7.6	
Naphthalene	1.078	1.121		4.0	
2-Fluorobiphenyl	1.502	1.618		7.7	
2,4,6-Tribromophenol	0.249	0.276		10.8	
Terphenyl-d14	1.112	1.118		0.5	

All other compounds must meet a minimum RRF of 0.010.