

Surrogate Summary

SW-846

SDG No.: Q3103

Client: ENTACT

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB169719BL	PB169719BL	2-Fluorophenol	150	119	79		15 (23)	110 (138)
		Phenol-d6	150	109	73		15 (10)	110 (134)
		Nitrobenzene-d5	100	73.4	73		30 (67)	130 (132)
		2-Fluorobiphenyl	100	75.1	75		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	111	74		15 (44)	110 (137)
		Terphenyl-d14	100	71.7	72		30 (42)	130 (152)
PB169719BS	PB169719BS	2-Fluorophenol	150	116	77		15 (23)	110 (138)
		Phenol-d6	150	109	73		15 (10)	110 (134)
		Nitrobenzene-d5	100	69.0	69		30 (67)	130 (132)
		2-Fluorobiphenyl	100	68.9	69		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	113	75		15 (44)	110 (137)
		Terphenyl-d14	100	72.6	73		30 (42)	130 (152)
PB169719BSD	PB169719BSD	2-Fluorophenol	150	116	78		15 (23)	110 (138)
		Phenol-d6	150	110	74		15 (10)	110 (134)
		Nitrobenzene-d5	100	69.8	70		30 (67)	130 (132)
		2-Fluorobiphenyl	100	69.9	70		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	113	76		15 (44)	110 (137)
		Terphenyl-d14	100	72.7	73		30 (42)	130 (152)
Q3103-01	TW-WTS-14	2-Fluorophenol	150	66.4	44		15 (23)	110 (138)
		Phenol-d6	150	37.4	25		15 (10)	110 (134)
		Nitrobenzene-d5	100	85.3	85		30 (67)	130 (132)
		2-Fluorobiphenyl	100	82.9	83		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	132	88		15 (44)	110 (137)
		Terphenyl-d14	100	78.0	78		30 (42)	130 (152)