

## Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3103

Analytical Method: 8270E

Client: ENTACT

DataFile: BP025774.D

| Lab Sample ID | Parameter              | Spike | Result | Unit | Rec | RPD | Qual | RPD  | Limits  |           | RPD |
|---------------|------------------------|-------|--------|------|-----|-----|------|------|---------|-----------|-----|
|               |                        |       |        |      |     |     |      | Qual | Low     | High      |     |
| PB169719BS    | Phenol                 | 50    | 38.7   | ug/L | 77  |     |      |      | 20 (66) | 160 (118) |     |
|               | 1,4-Dichlorobenzene    | 50    | 37.8   | ug/L | 76  |     |      |      | 70 (76) | 130 (103) |     |
|               | 1,2,4-Trichlorobenzene | 50    | 38.3   | ug/L | 77  |     |      |      | 70 (41) | 130 (115) |     |
|               | Naphthalene            | 50    | 37.5   | ug/L | 75  |     |      |      | 70 (64) | 130 (107) |     |