

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3103

Client ID: SSTDCCC040

Date Analyzed: 09/19/2025

Lab File ID: BP025792.D

Time Analyzed: 10:12

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

01

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	907902	17.172	942075	21.619	1041360	24.971
UPPER LIMIT	1815800	17.672	1884150	22.119	2082720	25.471
LOWER LIMIT	453951	16.672	471038	21.119	520680	24.471
EPA SAMPLE NO.						
TW-WTS-14	846153	17.18	903941	21.63	981310	25.00

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.