

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3103</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>ENTACT</u>	Datafile :	<u>VN087850.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0916WBS01	Methyl tert-butyl Ether	20	18.0	ug/L	90			70 (78)	130 (114)	
	Carbon Tetrachloride	20	19.2	ug/L	96			70 (77)	130 (113)	
	Chloroform	20	18.9	ug/L	95			70 (79)	130 (113)	
	1,1,1-Trichloroethane	20	18.7	ug/L	94			70 (80)	130 (108)	
	Benzene	20	19.1	ug/L	96			70 (82)	130 (109)	
	Toluene	20	18.6	ug/L	93			70 (82)	130 (110)	
	Tetrachloroethene	20	19.2	ug/L	96			70 (67)	130 (123)	
	Ethyl Benzene	20	18.2	ug/L	91			70 (83)	130 (109)	
	m/p-Xylenes	40	37.6	ug/L	94			70 (82)	130 (110)	
	o-Xylene	20	18.8	ug/L	94			70 (83)	130 (109)	