

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>GENV01</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3108</u>
Instrument ID:	<u>BNA_P</u>	Calibration Date/Time:	<u>09/19/2025 10:12</u>
Lab File ID:	<u>BP025792.D</u>	Init. Calib. Date(s):	<u>09/11/2025 09/11/2025</u>
EPA Sample No.:	<u>SSTDCCC040</u>	Init. Calib. Time(s):	<u>08:56 15:20</u>
GC Column:	<u>ZB-GR</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.240	1.228		-1.0	
Benzaldehyde	0.987	1.065		7.9	
Phenol-d6	1.648	1.562		-5.2	
bis(2-Chloroethyl)ether	1.396	1.332		-4.6	
2,2-oxybis(1-Chloropropane)	2.283	2.240		-1.9	
Acetophenone	0.534	0.523		-2.1	
n-Nitroso-di-n-propylamine	1.149	1.058	0.050	-7.9	
Nitrobenzene-d5	0.453	0.446		-1.5	
Hexachloroethane	0.602	0.587		-2.5	
Nitrobenzene	0.407	0.402		-1.2	
Isophorone	0.797	0.758		-4.9	
bis(2-Chloroethoxy)methane	0.461	0.460		-0.2	
Naphthalene	1.078	1.036		-3.9	
4-Chloroaniline	0.449	0.439		-2.2	
Hexachlorobutadiene	0.229	0.218		-4.8	20.0
Caprolactam	0.122	0.108		-11.5	
2-Methylnaphthalene	0.693	0.653		-5.8	
Hexachlorocyclopentadiene	0.330	0.324	0.050	-1.8	
2-Fluorobiphenyl	1.502	1.472		-2.0	
1,1-Biphenyl	1.468	1.478		0.7	
2-Chloronaphthalene	1.156	1.140		-1.4	
2-Nitroaniline	0.385	0.388		0.8	
Dimethylphthalate	1.527	1.459		-4.5	
Acenaphthylene	1.915	1.848		-3.5	
2,6-Dinitrotoluene	0.324	0.327		0.9	
3-Nitroaniline	0.340	0.339		-0.3	
Acenaphthene	1.104	1.057		-4.3	20.0
Dibenzofuran	1.800	1.711		-4.9	
2,4-Dinitrotoluene	0.460	0.457		-0.7	
Diethylphthalate	1.549	1.476		-4.7	
4-Chlorophenyl-phenylether	0.721	0.669		-7.2	
Fluorene	1.431	1.343		-6.2	
4-Nitroaniline	0.349	0.351		0.6	
n-Nitrosodiphenylamine	0.612	0.616		0.7	20.0
2,4,6-Tribromophenol	0.249	0.234		-6.0	
4-Bromophenyl-phenylether	0.220	0.220		0.0	
Hexachlorobenzene	0.245	0.236		-3.7	
Atrazine	0.234	0.230		-1.7	
Phenanthrene	1.135	1.108		-2.4	

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>GENV01</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3108</u>
Instrument ID:	<u>BNA_P</u>	Calibration Date/Time:	<u>09/19/2025 10:12</u>
Lab File ID:	<u>BP025792.D</u>	Init. Calib. Date(s):	<u>09/11/2025 09/11/2025</u>
EPA Sample No.:	<u>SSTDCCC040</u>	Init. Calib. Time(s):	<u>08:56 15:20</u>
GC Column:	<u>ZB-GR</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Anthracene	1.151	1.139		-1.0	
Carbazole	1.070	1.053		-1.6	
Di-n-butylphthalate	1.313	1.304		-0.7	
Fluoranthene	1.361	1.339		-1.6	20.0
Pyrene	1.335	1.325		-0.7	
Terphenyl-d14	1.112	1.080		-2.9	
Butylbenzylphthalate	0.585	0.598		2.2	
3,3-Dichlorobenzidine	0.473	0.476		0.6	
Chrysene	1.305	1.279		-2.0	
Bis(2-ethylhexyl)phthalate	0.856	0.864		0.9	
Di-n-octyl phthalate	1.523	1.524		0.1	20.0
Indeno(1,2,3-cd)pyrene	1.454	1.449		-0.3	
Dibenzo(a,h)anthracene	1.190	1.182		-0.7	
1,2,4,5-Tetrachlorobenzene	0.606	0.607		0.2	
1,4-Dioxane	0.520	0.520		0.0	20.0

All other compounds must meet a minimum RRF of 0.010.