

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3108

Client ID : SSTDCCC040

Date Analyzed: 09/19/2025

Lab File ID: BP025792.D

Time Analyzed: 10:12

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	182285	7.749	709982	10.51	456027	14.37
UPPER LIMIT	364570	8.249	1419960	11.013	912054	14.866
LOWER LIMIT	91142.5	7.249	354991	10.013	228014	13.866
EPA SAMPLE NO.						
01 MW2	165881	7.75	650117	10.52	426439	14.37

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.