

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3108

Client ID: SSTDCCC040

Date Analyzed: 09/19/2025

Lab File ID: BP025792.D

Time Analyzed: 10:12

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

01

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	907902	17.172	942075	21.619	1041360	24.971
UPPER LIMIT	1815800	17.672	1884150	22.119	2082720	25.471
LOWER LIMIT	453951	16.672	471038	21.119	520680	24.471
EPA SAMPLE NO.						
MW2	855270	17.18	903003	21.62	964980	24.98

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.