

Report of Analysis

Client:	G Environmental		Date Collected:	
Project:	61 Laurel Street		Date Received:	
Client Sample ID:	VN0916WBS01		SDG No.:	Q3108
Lab Sample ID:	VN0916WBS01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VN087850.D	1	09/16/25 15:25	VN091625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	18.5		0.22	1.00	ug/L
74-87-3	Chloromethane	19.0		0.32	1.00	ug/L
75-01-4	Vinyl Chloride	17.9		0.26	1.00	ug/L
74-83-9	Bromomethane	19.8		1.40	5.00	ug/L
75-00-3	Chloroethane	18.1		0.47	1.00	ug/L
75-69-4	Trichlorofluoromethane	19.4		0.33	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	18.3		0.25	1.00	ug/L
75-65-0	Tert butyl alcohol	90.0		5.50	25.0	ug/L
75-35-4	1,1-Dichloroethene	18.2		0.23	1.00	ug/L
67-64-1	Acetone	75.9		1.50	5.00	ug/L
75-15-0	Carbon Disulfide	19.4		0.21	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	18.0		0.16	1.00	ug/L
79-20-9	Methyl Acetate	19.6		0.27	1.00	ug/L
75-09-2	Methylene Chloride	19.2		0.28	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	18.3		0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	18.0		0.23	1.00	ug/L
110-82-7	Cyclohexane	17.7		1.50	5.00	ug/L
78-93-3	2-Butanone	87.0		0.98	5.00	ug/L
56-23-5	Carbon Tetrachloride	19.2		0.25	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	19.4		0.19	1.00	ug/L
74-97-5	Bromochloromethane	17.9		0.22	1.00	ug/L
67-66-3	Chloroform	18.9		0.25	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	18.7		0.20	1.00	ug/L
108-87-2	Methylcyclohexane	17.4		0.16	1.00	ug/L
71-43-2	Benzene	19.1		0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	18.1		0.22	1.00	ug/L
79-01-6	Trichloroethene	18.9		0.090	1.00	ug/L
78-87-5	1,2-Dichloropropane	18.4		0.20	1.00	ug/L
75-27-4	Bromodichloromethane	19.1		0.22	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	92.8		0.68	5.00	ug/L

Report of Analysis

Client:	G Environmental		Date Collected:	
Project:	61 Laurel Street		Date Received:	
Client Sample ID:	VN0916WBS01		SDG No.:	Q3108
Lab Sample ID:	VN0916WBS01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VN087850.D	1	09/16/25 15:25	VN091625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
108-88-3	Toluene	18.6		0.14	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	17.9		0.17	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	18.7		0.16	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	19.4		0.21	1.00	ug/L
591-78-6	2-Hexanone	90.1		0.89	5.00	ug/L
124-48-1	Dibromochloromethane	18.9		0.18	1.00	ug/L
106-93-4	1,2-Dibromoethane	19.0		0.15	1.00	ug/L
127-18-4	Tetrachloroethene	19.2		0.23	1.00	ug/L
108-90-7	Chlorobenzene	19.2		0.12	1.00	ug/L
100-41-4	Ethyl Benzene	18.2		0.13	1.00	ug/L
179601-23-1	m/p-Xylenes	37.6		0.24	2.00	ug/L
95-47-6	o-Xylene	18.8		0.12	1.00	ug/L
100-42-5	Styrene	18.8		0.15	1.00	ug/L
75-25-2	Bromoform	21.0		0.19	1.00	ug/L
98-82-8	Isopropylbenzene	16.7		0.12	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	17.5		0.26	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	18.0		0.16	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	18.2		0.19	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	17.9		0.16	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	16.4		0.53	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	18.0		0.20	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	17.8		0.20	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	44.1		74 - 125	88%	SPK: 50
1868-53-7	Dibromofluoromethane	47.5		75 - 124	95%	SPK: 50
2037-26-5	Toluene-d8	45.0		86 - 113	90%	SPK: 50
460-00-4	4-Bromofluorobenzene	46.8		77 - 121	94%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	212000	8.206			
540-36-3	1,4-Difluorobenzene	417000	9.082			
3114-55-4	Chlorobenzene-d5	385000	11.847			
3855-82-1	1,4-Dichlorobenzene-d4	210000	13.77			

Report of Analysis

Client:	G Environmental		Date Collected:	
Project:	61 Laurel Street		Date Received:	
Client Sample ID:	VN0916WBS01		SDG No.:	Q3108
Lab Sample ID:	VN0916WBS01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VN087850.D	1	09/16/25 15:25	VN091625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products