

Surrogate Summary

SDG No.: **Q3108**

Client: **G Environmental**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3108-01	MW2	1,2-Dichloroethane-d4	50	48.2	96		74	125
		Dibromofluoromethane	50	50.5	101		75	124
		Toluene-d8	50	48.9	98		86	113
		4-Bromofluorobenzene	50	50.2	100		77	121
VN0916WBL01	VN0916WBL01	1,2-Dichloroethane-d4	50	49.1	98		74	125
		Dibromofluoromethane	50	50.7	101		75	124
		Toluene-d8	50	47.1	94		86	113
		4-Bromofluorobenzene	50	45.7	91		77	121
VN0916WBS01	VN0916WBS01	1,2-Dichloroethane-d4	50	44.1	88		74	125
		Dibromofluoromethane	50	47.5	95		75	124
		Toluene-d8	50	45.0	90		86	113
		4-Bromofluorobenzene	50	46.8	94		77	121
VN0916WBSD0	VN0916WBSD01	1,2-Dichloroethane-d4	50	43.5	87		74	125
		Dibromofluoromethane	50	48.6	97		75	124
		Toluene-d8	50	45.6	91		86	113
		4-Bromofluorobenzene	50	46.4	93		77	121